

SW-9001 Shoulder Wrap

***CAUTION: DO NOT OVERHEAT!!** Due to the nature of the gel, it may only feel warm to the touch, but if the wrap has been overheated, too much heat may be transferred to the body and can cause burns.

DIRECTIONS FOR USE

COLD THERAPY:


Store the wrap in the freezer for a minimum of two (2) hours. It may be left in the freezer indefinitely until ready to use. It will remain flexible even at -20°F. Remove the wrap prior to defrosting the freezer. **Note: Always store product in a plastic bag while in freezer.**

HEAT THERAPY:

The most convenient method of heating the wrap is in a **microwave oven**. Suggested heating times: **600-700 watt microwave:** 60-90 Seconds. Flip the wrap to the opposite side and heat for 60-90 Seconds. Then refold the wrap (inside to out) and heat for 45 Seconds. **1100 watt:** 50 Seconds per side. **Note: all microwaves vary in heating and may have "hot spots". Be cautious when heating.**

Wait one minute (60 seconds) before using. The wrap should feel **WARM** to the touch, **NOT HOT**. Hold the wrap firmly against the skin for 30 seconds **before re-heating***. If the wrap is not warm enough, heat in 20 second increments and rotate the product each time.

The wrap can also be heated in a **conventional oven** at 250°F for 15-20 minutes. Place wrap on cookie sheet or baking dish in the center of the oven. Do not allow the straps or buckles to touch the heating element. Rotate the wrap after 6 minutes, top to bottom and if folded, inside to out.



How to apply:

1. The SW9001 Shoulder Wrap is universal for either right or left shoulder. Lay the pad out flat first.
2. Insert Male Buckle B into Female Buckle B.
3. Place the wide side of the wrap on the desired shoulder with the narrow side laying over the bicep area.
4. Bring Strap A under the opposite arm and Insert Male Buckle C into Female Buckle C. Note: Strap A is Adjustable.
5. Loop Strap D through buckle D and secure strap D back onto itself.
6. Wrap the Narrow part of the Shoulder Wrap around the bicep area and secure Strap E around bicep. Strap E will hook onto itself.

DO NOT GET THE PRODUCTS WET:

The gel will absorb water and swell. If the product becomes damp, lay it on a flat surface and allow it to fully air dry before continuing use. Do not use wet towels under the gel.

CLEANING INSTRUCTIONS:

Remove the outside cover and hand wash in cold water and mild soap. Hang up to dry. Wipe the white inner cover with a damp (not wet) cloth using mild soap or disinfectant. Allow the product to fully air dry before continuing to use.

Treating the World Well[®]

southwest technologies inc.

1746 Levee Road, North Kansas City, MO 64116

phone: (800) 247-9951 phone: (816) 221-2442 fax: (816) 221-3995

email: info@elastogel.com • website: www.elastogel.com



E C Rep Ltd
5 Fitzwilliam Square East
Dublin 2, Ireland
Tel: +353 1 2 544 944
info@ecrep.ie
www.ecrep.ie